Dutscher

ClearLine®

TECHNICAL DATASHEET

PART NUMBER 713114

ISO 8655

The ISO8655 norm guarantees that the tips used do not alter the technical characterics of the pipettes, as given by the manufacturer.

IVD

Products are certified in accordance with the 79/98/EC directive related to In Vitro Diagnostic medical devices.

ClearLine Tips

Cat. N°	713114
Volume (µl)	10 µl
Low binding	Yes
Filter	Yes
Sterile	Yes (e-beam)
Autoclavable	No
Material	Medical grade virgin polypropylene
Packaging	On hinged-lid rack
Storage	Store in a clean, dry location at ambiant temperature
Quality Control	Each lot certified : RNase, DNase-free, Human DNA free, Endotoxin & pyrogen free
Traceability	On each rack : 5-digit lot number The tip's volume & Cat. N° Expiry date



Units :

960 tips

Dominique Dutscher SAS 30 rue de l'Industrie ZAC BRUMATH NORD BP62 67172 BRUMATH CEDEX
Tél : +33 (0)3 88 59 33 90 Fax : +33 (0)3 88 59 33 99 Mail : info@dutscher.com Web : www.dutscher.com